## **CAUTION: BEDSIDE MISCONNECTION BISKS**

The Gabriel® Feeding Tube with Balloon has a caution label attached at its proximal end stating "Enteral Feed/Use Only, MR Conditional" along with instructions on deflating the balloon before removal. This label is to alert staff that the Gabriel Feeding Tube with Balloon is for enteral use only and to be aware of potential bedside misconnections. Bedside misconnections can cause serious iniury to the patient. The connectors listed below are known to potentially engage and connect to the Gabriel Feeding Tube with Balloon. Please ensure that connections with the following and similar items are not made:

- Female Luer Lock with Wings (.229" minor diameter and .304" across tabs)
- Nebulizer Oxygen Port Foley Balloon Cuff Port
- Christmas Tree Oxygen Connector Tracheotomy Cuff Port
- Endo Tracheal Tube cuff port
  Corrugated Oxygen Nipple

This Gabriel Feeding Tube's connector is intended to connect to enteral feeding pump lines, enteral feeding gravity set lines, and the Lopez 3-way Valve (ICU Medical K915171).

## **PRECAUTIONS**

- Tube balloon (3 ml) and pilot balloon (3 ml) must be deflated before removal from patient. Please remove all 6 ml.
- Please read all illustrated instructions for the procedure prior to use.
- For technical questions please call (478) 335-8311.

## **MRI SAFETY INFORMATION**

Non-clinical testing has demonstrated that the Gabriel Feeding Tube with Balloon Size 12 Fr is MR Conditional. A patient with this device can be safely scanned in an MR system meeting the following conditions:

- Static magnetic field of 1.5-Tesla (1.5 T) or 3-Tesla (3 T).
- Maximum spatial field gradient of 3,000 G/cm (30.0 T/m).
- Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 4.0 W/kg.

RF Heating - Under the scan conditions defined above, the Gabriel Feeding Tube with Balloon Size 12 Fr is expected to produce a maximum temperature rise of less than 5.0 °C after 15 minutes of continuous scanning. Caution: The RF heating behavior does not scale with static field strength. Devices that do not exhibit detectable heating at one field strength may exhibit high values of localized heating at another field strength. The breaking of the tube and exposure of the coil could lead to greater rise in RF heating during MRI scan.

MR Artifact - In non-clinical testing, the image artifact caused by the device extends approximately 14 mm from the Gabriel Feeding Tube with Balloon Size 12 Fr when imaged with a gradient-echo pulse sequence in a 3 T MRI system.


Expiration date: November 2023 Print date: November 2019

U.S. Patent #9,713,578 Japan Patent #6357311 **China Patent** #ZL201310710594.0 **European Patent** #2 745 828

CE 0197



©2019 | Printed in China | GFTB-IFU 412-12Fr 111319



## **CONTENTS**

- Gabriel Feeding Tube with Balloon, preassembled with stylet
- Cotton swab
  Numbing gel
  CO2 Sampling Line
- Lubricating gel Silk suture thread
- 10 cc Luer Lock syringe Skin adhesive
- Butterfly-tape to secure tube at end of procedure

## ADDITIONAL MATERIALS NEEDED

1 Cup / Gloves / 2x2 Gauze

## **RE-ORDER INFORMATIONS**

Size	Length	<b>Balloon Size</b>	Reorder #
12 Fr	130 cm	3 ml	GFTB 412
10 Fr	100 cm	3 ml	GFTB 410
8 Fr	80 cm	1 ml	GFTB 408

Design and specification developed by: Syncro Medical Innovations, Inc. 515 Mulberry Street, Suite 200 Macon, Georgia 31201, USA gabrielfeedingtube.com / Tel: 888-980-9159 / Fax: (716) 806-3504

# GABRIEL FEEDING TUBE WITH BALLOON

Preassembeled with Stylet

- Radio-opaque and disposable (Intended for single-patient use only, non-sterile tube)
- Purple colored proximal Y shaped connector
- Purple colored 3 ml pilot balloon at proximal end and 3 ml inflatable balloon at distal end.
- Single line mark at 50 cm (Corresponds to placement at lower end of esophagus)
- Double line mark at 80 cm (Corresponds to placement at duodenal bulb)
- Triple line mark at 110 cm (Corresponds to distal duodenal placement)
- Preassembled with stylet
- Read all precautions and instructions for use before use.

**MR conditional.** (See last page for MRI safety information.)

**Caution:** Federal law restricts this device to sale by or on the order of a physician


Indications for Use: The Gabriel Feeding Tube with balloon functions as a conduit to facilitate enteral feeding and may be used in adult or elderly patients who cannot consume an adequate diet orally. Small bowel feeding may be indicated for patients with functioning gut who require short to moderate term feeding support such as post-trauma patients, burn patients, general trauma patients, high-risk patients prone to tube misplacement complications, and patients in whom malnutrition exists—or may result—secondary to an underlying disease or condition.

MADE IN CHINA


Syncro Medical Innovations, Inc. 515 Mulberry Street, Suite 200 Macon, Georgia 31201, USA




## 888-980-9159 gabrielfeedingtube.com




- Connect CO<sub>2</sub> Sampling Line to Stylet End Hub and cap the side port.
- Apply provided numbing gel to nostril using the cotton-tipped swab provided. Wait two minutes.
- Insert silk thread into the distal end holes. • Apply lubricating gel to the tube distal tip and nostril.




- Hold both ends of silk thread at six o'clock position and insert the tube toward the back of the head to the nasopharynx.
- · Pull both ends of thread to flex the tip and guide the tube downward to the oropharynx.
- Remove the thread completely. (The use of thread steering technique is optional.)




- When the 18 cm mark is at the nostril, ask the patient to swallow in order to advance the tube into the esophagus.
- When the 30 cm mark is at the nostril, the tube tip should be in mid-esophagus. Cough should be absent in conscious patients.







- · Once esophageal placement is confirmed, continue advancing the tube down the esophagus and into the stomach.
- When the tube reaches the 70 cm mark. pull the stylet out gradually as you advance more tube.



- Continue advancing the tube until the 100 cm mark is at the nostril then remove the stylet completely.
- Apply skin prep-site adhesive solution to the nose, then secure the tube at 100 cm using the nasal butterfly-tape provided.



Inject 60 cc of warm water to stimulate gastric peristalsis to help the tube migrate deeper into the small intestine.



Obtain abdominal x-ray to confirm gastric or intestinal placement before initiating feeding. X-ray is the gold standard to confirm final tube placement. Tube migration can be monitored by abdominal x-ray as needed. X-rays above show (on left) tube in the stomach shortly after placement and (on right) the tube in ieiunum 12 hours after insertion.



Sedated patients who are unresponsive to trachea misplacement should be connected to pulseoximetery and CO<sub>2</sub> monitor compatible with the provided CO2 sampling line. Using the Luer Lock syringe, inflate the tube distal end balloon and pilot balloon with 6 cc of air and observe end-tidal CO2 wave and pulse-oximetry. Detection of endtidal CO<sub>2</sub> wave or drop in pulse-oximetery by 5 or more points indicates misplacement in the trachea. Deflate balloon and withdraw to 18 cm mark then reinsert into esophagus.

Optional: Tube balloons can be inflated with 6 cc water and patient turned on right side down to facilitate distal tube placement.



- Initiate feeding and maintain as tolerated by the patient.
- Deflate tube balloons after 48 hours or before removal.
- Precaution: Do not use with feeding pumps that can generate pressure greater than 40 psi.
- Flush tube with warm water (20-30 ml) prior to and after administering medications, every 3-4 hours during continuous feeding, and after intermittent feedings.
- Precaution: Do not use a syringe smaller than 30 cc when irrigating the feeding tube.

